

24. Medical physics

Starter Quiz · 课前小测

A-Level Physics

Name: _____ Score: _____

1. A piezo-electric crystal:

- ☐ changes shape with a p.d., and makes an e.m.f. when squeezed
- ☐ only emits light
- ☐ stores charge like a battery
- ☐ conducts current in one direction only

2. X-rays are produced when fast electrons:

- ☐ hit a metal target and slow down sharply
- ☐ pass through a magnet
- ☐ are heated gently
- ☐ travel through glass

3. A PET tracer is:

- ☐ a positron (β^+) emitter taken up by active tissue
- ☐ a strong magnet
- ☐ an X-ray source
- ☐ an ultrasound gel

4. An echo returns after 2.0×10^{-4} s; the speed of sound is $1500 \frac{\text{m}}{\text{s}}$. How deep is the boundary?
_____ m

5. Most of the electrons' kinetic energy in an X-ray tube becomes heat, not X-rays.

- ☐ True ☐ False

6. Why is a tracer's half-life chosen as a compromise? Select **all** that apply.

- ☐ short enough that the patient's dose after the scan is small
- ☐ short enough that the activity is high during the scan
- ☐ long enough that it has not decayed before reaching the tissue
- ☐ long enough to stay in the body for later scans

Tick every correct option.

7. An ultrasound pulse returns 60 microseconds after transmission, with a speed of sound of 1500 m/s. How deep is the boundary, in cm?

8. The minimum X-ray wavelength from a tube at accelerating voltage V is:

- ☐ $\frac{hc}{eV}$
- ☐ $\frac{eV}{hc}$
- ☐ $\frac{h}{eV}$
- ☐ $hceV$

9. When the emitted positron meets an electron, they:

- ☐ annihilate into two photons
- ☐ bounce off each other
- ☐ form a stable atom
- ☐ do nothing

10. A higher ultrasound frequency has which effects? Select **all** that apply.

- ☐ a shorter wavelength
- ☐ better resolution
- ☐ stronger attenuation, so less penetration
- ☐ a larger reflection at every boundary

Tick every correct option.

1. **changes shape with a p.d., and makes an e.m.f. when squeezed**

Both effects let one crystal send ultrasound (apply p.d.) and detect it (read the e.m.f.).

2. **hit a metal target and slow down sharply**

The sudden deceleration on the target gives out X-ray photons (Bremsstrahlung).

3. **a positron (β^+) emitter taken up by active tissue**

It is a β^+ emitter on a molecule (like glucose) that active tissue absorbs more of.

4. **0.15 m**

$$d = \frac{ct}{2} = \frac{1500 \times 2.0 \times 10^{-4}}{2} = 0.15 \text{ m (the pulse travels there and back).}$$

5. **True**

Only a small fraction becomes X-rays — which is why the target must be cooled.

6. **short enough that the patient's dose after the scan is small; short enough that the activity is high during the scan; long enough that it has not decayed before reaching the tissue**

Fluorine-18 has 110 minutes; oxygen-15 has only 2 and must be made on the spot. Leaving activity in the patient afterwards is the thing to avoid, not to aim for.

7. **4.5**

$d = ct/2 = 4.5 \text{ cm}$. Halve it: the pulse travelled there AND back. Forgetting the 2 gives 9.0 cm.

8. $\frac{hc}{eV}$

From $hf_{\max} = eV$ and $c = f\lambda$: $\lambda_{\min} = \frac{hc}{eV}$.

9. **annihilate into two photons**

Matter and antimatter annihilate: their mass turns into two photons.

10. **a shorter wavelength; better resolution; stronger attenuation, so less penetration**

Reflection is set by the impedance mismatch, not the frequency. The frequency choice is a resolution-against-penetration compromise, so a deep organ is scanned at a lower frequency.